

Ribospin™ vRD

VIRAL RNA/DNA PURIFICATION HANDBOOK

REF

302-150/302-103

H B

HB3200



GeneAll Bldg., 303-7, Dongnamro, Songpa-gu,
Seoul, Korea 05729

Customer & Technical Support

Should you have any further questions, do not hesitate to contact us.

We appreciate your comments and advice.

Contact Information

www.geneall.com

Tel : 82-2-407-0096

Fax : 82-2-407-0779

E-mail (Order/Sales) : sales@geneall.com

E-mail (Tech. Info.) : tech@geneall.com

Visit GeneAll® Community

www.geneall.com

Trademarks

Hybrid-Q™, Exprep™, Exfection™, Expin™, Exgene™, GenEx™, DirEx™, Hybrid-R™, RiboEx™, Ribospin™, Riboclear™, Allspin™, RiboSaver™, EzClear™, EzSep™, EzPure™, EzPass™, AmpONE™, AmpMaster™, RealAmp™, HyperScript™, ProtinEx™, PAGESTA™, STEADi™, GENTi™, SolMate™ are trademarks of GeneAll Biotechnology Co., Ltd.

© 2021 GeneAll Biotechnology, all rights reserved.

This protocol handbook is included in :

GeneAll® Ribospin™ vRD (302-I 50, 302-I 03)

Visit www.geneall.com for FAQ, Q&A and more information.

Used symbols and Markings

	Catalog number		In vitro diagnostic medical device
	Batch number		Handbook code
	Use by		Consult instruction for use
	Manufacturer information		Contains sufficient for <N> tests
	Do not reuse		Temperature limitation
	Production date		European Authorized Representative
	Important note		Contains the concentrated solution. Additional material must be added before use
	Write down the current date after adding ethanol to the bottle		Mark up after adding ethanol

INDEX

04	Index
05	Kit Contents
	Product Specifications
06	Quality Control
	Storage Conditions
	Safety Information
	Preventing RNase Contamination
07	Product Description
08	Protocol
10	Troubleshooting Guide
12	Ordering Information

Kit Contents

Components	Quantity		Storage
	302-150	302-103	
Cat. No.			
No. of preparation	50	300	 Room temperature (15~25°C)
Buffer VL	30 ml	170 ml	
Buffer RB1 (concentrate) *	8 ml	48 ml	
Buffer RBW (concentrate) *	13 ml	77 ml	
Buffer RNW (concentrate) * †	6 ml	34 ml	
Nuclease-free water	15 ml	20 ml	
Column Type V (mini) (with collection tube)	50	300	
1.5 ml microcentrifuge tube	50	300	
Protocol Handbook	1	1	



* Before first use, add absolute ethanol (ACS grade or better) into Buffer RB1, RBW and RNW as indicated on the bottle.

† Contains sodium azide as a preservative

Product Specifications

Ribospin™ vRD	
Type	Using spin column
Maximum volume of starting samples	300 µl/prep
Preparation time	20 min
Maximum loading volume	800 µl
Minimum elution volume	30 µl

Quality Control

All components of GeneAll® Ribospin™ vRD are manufactured in strictly clean conditions, and their degree of cleanliness is monitored periodically.

To maintain consistency, a quality control process is carried out thoroughly from lot to lot and only the qualified kits are approved for delivery according to ISO 9001:2008 and EN ISO 13485:2012.

Storage Conditions

All components of GeneAll® Ribospin™ vRD should be stored at room temperature (15~25°C).

During shipment or storage under cool ambient condition, a precipitate can form in Buffer VL. In such a case, heat the bottle to 56°C to dissolve completely. GeneAll® Ribospin™ vRD is guaranteed until the expiration data printed on the product box.

Safety Information

The buffers included in the GeneAll® Ribospin™ vRD contain irritants which are harmful when in contact with skin or eyes, or when inhaled or swallowed. Care should be taken when handling such materials. Always wear gloves and eye protection, and follow standard safety precautions.



Buffer VL, RBI, and RBW contain chaotropes, which can form highly reactive compounds when combined with bleach. Do NOT add bleach or acidic solutions directly to the sample-preparation waste.

Preventing RNase contamination

RNase can be introduced accidentally during RNA purification. Wear disposable gloves always, because skin often contains bacteria and molds that can be a source of RNase contamination. Use sterile, disposable plastic wares and automatic pipettes to prevent cross-contamination of RNase from shared equipment.

Product Description

GeneAll® Ribospin™ vRD provides a convenient method for isolation of RNA and DNA from cell-free fluid, cell-culture medium, plasma, serum, swab, urine and virus-infected samples.

GeneAll® Ribospin™ vRD utilizes the glass fiber membrane technology for the fastest and the most convenient nucleic acid isolation as a sufficient level for downstream application instead of conventional alcohol precipitation or phenol/chloroform extraction.

The buffer system of Ribospin™ vRD provides the effective binding condition of RNA and DNA to glass fiber membrane and the impurities on the membrane are washed away by two different wash buffers. At least, pure RNA and DNA are eluted in Nuclease-free water. The whole procedure may take only 15 minutes at room temperature and the eluate is suitable for PCR, RT-PCR, or any downstream application without further manipulation. The purified nucleic acid should be treated with care because RNA is very sensitive to contaminants such as RNases, often found on general lab ware and dust. To ensure RNA-stability after extraction, it is recommended to store at 4°C for immediate analysis or to freeze at -70°C for long-term storage.

PROTOCOL FOR

Ribospin™ vRD

Equipment and reagents to be supplied by user

- * Ethanol (>99%, ACS grade or better)
- * 1.5 ml microcentrifuge tubes
- * Micropipettes and sterile pipet tips
- * Centrifuge capable of attaining 10,000 x g
- * Vortex mixer



- Ethanol (>99%, ACS grade or better) must be added before the first use of Buffer RB I, RBW and RNW. Please refer to the information on the label of each bottle.
- If a precipitate is formed in Buffer VL, heat to 56°C to dissolve completely before use.

- 1. Transfer up to 300 µl sample (swab-storage media, cell-free fluid, cell-culture supernatant, plasma, serum, urine) in 1.5 ml microcentrifuge tube.**
- 2. Add 500 µl Buffer VL to the tube and lyse the sample by pipetting or vortexing.**

The volume of Buffer VL can be adjusted in proportion to the volume of sample.
For proper lysis, the complete mixing of sample and Buffer VL is essential.
- 3. Incubate the lysate for 10 min at room temperature.**

After this step, briefly centrifuge the tube to remove drops from the inside of the lid.
- 4. Add 700 µl Buffer RB I to the lysate and mix thoroughly by inverting or vortexing.**

The volume of Buffer RB I can be adjusted in proportion to the volume of lysate.
* Do NOT centrifuge at this step.
- 5. Transfer up to 750 µl of the mixture to a Column Type V (mini).**

6. Centrifuge at $\geq 10,000 \times g$ for 30 sec at room temperature.

Discard the pass-through and reinsert the mini column back into the same tube.

7. Repeat step 5~6 with the remainder of the sample.

Discard the pass-through and reinsert the mini column back into the same tube.

8. Add 500 μ l Buffer RBW to the mini column.

9. Centrifuge at $\geq 10,000 \times g$ for 30 sec at room temperature.

Discard the pass-through and reinsert the mini column back into the same tube.

10. Add 500 μ l Buffer RNW to the mini column.

11. Centrifuge at $\geq 10,000 \times g$ for 30 sec at room temperature.

Discard the pass-through and reinsert the mini column back into the same tube.

12. Centrifuge at $\geq 10,000 \times g$ for an additional 1 min at room temperature to remove residual wash buffer. Transfer the mini column to a new 1.5 ml microcentrifuge tube (provided).

Residual ethanol may interfere with downstream reactions. Care must be taken at this step for eliminating the carry-over Buffer RNW.

If the carry-over Buffer RNW still occurs, centrifuge again for 1 min at full speed before transferring the column to the new 1.5 ml microcentrifuge tube.

13. Add 30~50 μ l of Nuclease-free water to the center of the membrane in the mini column.

Let it stand for 1 min.

14. Centrifuge at $\geq 10,000 \times g$ for 1 min at room temperature.

Purified nucleic acid can be stored at 4°C for immediate analysis or at -70°C for long-term storage.

■ Troubleshooting Guide

Facts	Possible Causes	Suggestions
Low yield	Poor quality of starting material	Fresh sample or well-conserved sample should be used for good result. Repeated freezing and thawing the sample should be avoided.
	Low concentration of viral particle in the starting sample	Use more starting sample. If the amount of sample is more than 300 μ l, concentrate the volume to 300 μ l using a micro-concentrator.
	Inefficient or insufficient lysis	Be sure to incubate for 10 minutes at room temperature after adding Buffer VL. For proper lysis, the complete mixing of the sample and Buffer VL is essential.
	Improper elution	Add Nuclease-free water to the center of the mini column membrane and perform incubation for 1 minute before centrifugation.
	Precipitate in Buffer VL	A precipitate can be formed in Buffer VL at cool ambient temperature. It is because the Buffer VL is saturated and its solubility would be reduced at low temperature. Before experiment, any precipitate in the Buffer VL should be dissolved completely by heating the buffer at 56°C or above until it disappears.
	Degradation of RNA	RNase can be introduced during purification of nucleic acid. Be certain not to introduce any RNases during the procedure of later handling. Keep tubes closed whenever possible during the extraction and use RNase-free products with sterile and disposable plastic ware.
	Buffer RBI, RBW, or RNW was prepared incorrectly	Check that the concentrated Buffer RBI, RBW, and RNW were diluted with the correct volume of absolute ethanol.

Troubleshooting Guide

Facts	Possible Causes	Suggestions
Purified nucleic acid does not perform well in downstream application	Residual ethanol from Buffer RNW remains in eluate	Care must be taken for eliminating the carry-over Buffer RNW before elution step. The membrane of mini column should be kept completely dry via additional centrifugation (Step 12, page 9) or air-drying.
	Incorrect order of Buffer RBW and RNW	Ensure that Buffer RBW and RNW are used in the correct order during extraction. If used in the wrong order, perform the last washing step with Buffer RNW.

Ordering Information

Products	Scale	Size	Cat. No.	Type
----------	-------	------	----------	------

GeneAll® Hybrid-Q™ for rapid preparation of plasmid DNA

Plasmid Rapidprep	mini	50	100-150	spin
		200	100-102	

GeneAll® Exprep™ for preparation of plasmid DNA

Plasmid SV	mini	50	101-150	spin /
		200	101-102	vacuum
	Midi	26	101-226	spin /
		50	101-250	vacuum
		100	101-201	

GeneAll® Exfection™ for preparation of transfection-grade plasmid DNA

Plasmid LE (Low Endotoxin)	mini	50	111-150	spin /
		200	111-102	vacuum
	Midi	26	111-226	spin /
		100	111-201	vacuum
Plasmid EF (Endotoxin Free)	Midi	20	121-220	spin
		100	121-201	

GeneAll® Expin™ for purification of fragment DNA

Gel SV	mini	50	102-150	spin /
		200	102-102	vacuum
PCR SV	mini	50	103-150	spin /
		200	103-102	vacuum
CleanUp SV	mini	50	113-150	spin /
		200	113-102	vacuum
Combo GP	mini	50	112-150	spin /
		200	112-102	vacuum

GeneAll® Exgene™ for isolation of total DNA

Tissue SV	mini	100	104-101	spin /
		250	104-152	vacuum
	Midi	26	104-226	spin /
		100	104-201	vacuum
	MAXI	10	104-310	spin /
		26	104-326	vacuum
Tissue plus! SV	mini	100	109-101	spin /
		250	109-152	vacuum
	Midi	26	109-226	spin /
		100	109-201	vacuum
	MAXI	10	109-310	spin /
		26	109-326	vacuum

Products	Scale	Size	Cat. No.	Type
----------	-------	------	----------	------

GeneAll® Exgene™ for isolation of total DNA

Blood SV	mini	100	105-101	spin /
		250	105-152	vacuum
	Midi	26	105-226	spin /
		100	105-201	vacuum
	MAXI	10	105-310	spin /
		26	105-326	vacuum
Cell SV	mini	100	106-101	spin /
		250	106-152	vacuum
	MAXI	10	106-310	spin /
		26	106-326	vacuum
Clinic SV	mini	100	108-101	spin /
		250	108-152	vacuum
	Midi	26	108-226	spin /
		100	108-201	vacuum
	MAXI	10	108-310	spin /
		26	108-326	vacuum
Genomic DNA micro		50	118-050	spin
Plant SV	mini	100	117-101	spin /
		250	117-152	vacuum
	Midi	26	117-226	spin /
		100	117-201	vacuum
	MAXI	10	117-310	spin /
		26	117-326	vacuum
Soil DNA mini	mini	50	114-150	spin
Stool DNA mini	mini	50	115-150	spin
Viral DNA / RNA	mini	50	128-150	spin
FFPE Tissue DNA	mini	50	138-150	spin
		250	138-152	

GeneAll® GenEx™ for isolation of total DNA without spin column

GenEx™ Blood	Sx	100	220-101	solution
		500	220-105	
	Lx	100	220-301	solution
GenEx™ Cell	Sx	100	221-101	solution
		500	221-105	
	Lx	100	221-301	solution
GenEx™ Tissue	Sx	100	222-101	solution
		500	222-105	
	Lx	100	222-301	solution

Products	Scale	Size	Cat. No.	Type
----------	-------	------	----------	------

GeneAll® GenEx™ for isolation of total DNA

GenEx™ Plant	Sx	100	227-101	solution
	Mx	100	227-201	
	Lx	100	227-301	
GenEx™ Plant plus!	Sx	100	228-101	solution
	Mx	50	228-250	
	Lx	20	228-320	

GeneAll® DirEx™ series for preparation of PCR-template without extraction

DirEx™		100	250-101	solution
DirEx™ Fast-Tissue	96 T		260-011	solution
DirEx™ Fast-Cultured cell	96 T		260-021	solution
DirEx™ Fast-Whole blood	96 T		260-031	solution
DirEx™ Fast-Blood stain	96 T		260-041	solution
DirEx™ Fast-Hair	96 T		260-051	solution
DirEx™ Fast-Buccal swab	96 T		260-061	solution
DirEx™ Fast-Cigarette	96 T		260-071	solution

GeneAll® RNA series for preparation of total RNA

RiboEx™	mini	100	301-001	solution
		200	301-002	
Hybrid-R™	mini	100	305-101	spin
Hybrid-R™ Blood RNA mini		50	315-150	spin
Hybrid-R™ miRNA	mini	50	325-150	spin
RiboEx™ LS	mini	100	302-001	solution
		200	302-002	
Riboclear™	mini	50	303-150	spin
Riboclear™ plus!	mini	50	313-150	spin
Ribospin™	mini	50	304-150	spin
Ribospin™ II	mini	50	314-150	spin
		300	314-103	
Ribospin™ vRD	mini	50	302-150	spin
Ribospin™ vRD plus!	mini	50	312-150	spin
Ribospin™ vRD II	mini	50	322-150	spin
Ribospin™ Plant	mini	50	307-150	spin
Ribospin™ Seed / Fruit	mini	50	317-150	spin
Allspin™	mini	50	306-150	spin
RiboSaver™	mini	100	351-001	solution

Products	Scale	Size	Cat. No.	Type
----------	-------	------	----------	------

GeneAll® AmpONE™ for PCR amplification

Taq DNA polymerase		250 U	501-025	(2.5 U/μl)
		500 U	501-050	
		1,000 U	501-100	
Taq Premix	96 tubes	20 μl	526-200	solution
		50 μl	526-500	

GeneAll® AmpMaster™ for PCR amplification

Taq Master mix	0.5 ml x 2 tubes	541-010	solution
	0.5 ml x 10 tubes	541-050	solution

GeneAll® HyperScript™ for Reverse Transcription

Reverse Transcriptase	10,000 U	601-100	solution
RT Master mix	0.5 ml x 2 tubes	601-710	solution
One-step RT-PCR Master mix	0.5 ml x 2 tubes	602-110	solution
One-step RT-PCR Premix	96 tubes, 20 μl	602-102	solution

GeneAll® RealAmp™ for qPCR amplification

SYBR qPCR Master mix (2X, Low ROX)	200 rxn	20 μl	801-020	solution
	500 rxn	20 μl	801-050	
SYBR qPCR Master mix (2X, High ROX)	200 rxn	20 μl	801-021	solution
	500 rxn	20 μl	801-051	

Products	Size	Cat. No.	Type
----------	------	----------	------

GeneAll® Protein series

ProteinEx™ Animal cell / tissue	100 ml	701-001	solution
PAGESTA™ Reducing 5X SDS-PAGE Sample Buffer	1 ml × 10 tubes	751-001	solution

GeneAll® STEADI™ for automatic nucleic acid purification

I 2 Instrument		GST012	system
24 Instrument		GST024	system
Genomic DNA Cell / Tissue	96	401-104	kit
Genomic DNA Blood	96	402-105	kit
Total RNA	96	404-304	kit
Viral DNA / RNA	96	405-322	kit
CFC Seed DNA / RNA	96	406-C02	kit
Genomic DNA Plant	96	407-117	kit
Soil DNA	96	408-114	kit

GeneAll® GENTi™ 32 Ultimately flexible automatic extraction system

Automatic extrantion equipment		GTI032	system
Genomic DNA	48	901-048	strip
	96	901-096	plate
Viral DNA / RNA	48	902-048	strip
	96	902-096	plate
Whole Blood Genomic DNA	48	903-048	strip
	96	903-096	plate

Products	Scale	Size	Cat. No.	Type
----------	-------	------	----------	------

GeneAll® GENTi™ 32 Ultimately flexible automatic extraction system

Automatic extrantion equipment		GTI032A	system
Genomic DNA	48	901-048A	strip
	96	901-096A	plate
Viral DNA / RNA	48	902-048A	strip
	96	902-096A	plate
CFC Plant Pathogen	48	904-C01A	strip
	96	904-C02A	plate
Plant	48	904-048A	strip
	96	904-096A	plate

NOTE



GENEALL BIOTECHNOLOGY CO., LTD

GeneAll Bldg., 303-7, Dongnamro, Songpa-gu, Seoul, Korea 05729

E-mail : sales@geneall.com

Tel. 82-2-407-0096 Fax. 82-2-407-0779

www.geneall.com

Manufacturer site

A-1201 ~A-1204, Hanam Techno Valley UI Center,
947, Hanam-daero, Hanam-si, Gyeonggi-do, 12982, Korea

EC	REP
----	-----

MT Promedt Consulting GmbH
Altenhofstrasse 80, 66386 St. Ingbert,
Germany

Tel. +49 (0) 6894 581020

Fax. +49 (0) 6894 581021